



財團法人全國認證基金會  
Taiwan Accreditation Foundation

## Certificate of Accreditation

(Certificate No : L1735-240222)

This is to certify that

**Measurement Technology Co.,Ltd**  
**Calibration Laboratory of Southern Region Service Department**  
No.92, Dashe road, Dashe District, Kaohsiung city, Taiwan, (R.O.C.)

**is accredited in respect of laboratory**

**Accreditation Criteria** : ISO/IEC 17025:2017 ; CNS 17025:2018

**Accreditation Number** : 1735

**Originally Accredited** : February 16, 2007

**Effective Period** : February 16, 2022 to February 15, 2025

**Accredited Scope** : Calibration Field, see described in the Appendix



Scan to verify

*Yi-Ling Chen*

Yi-Ling Chen  
President, Taiwan Accreditation Foundation  
February 22, 2024

Accreditation Number : 1735  
 Laboratory Head : KAO, Wei-Chuan

calibration items	working standard brand /model	calibration method document name /no.	measurand level or range			measurement conditions /independent variable		smallest uncertainty	
			minimum value	units	maximum value	units	explanation		value
KA1001 Gauge Block (Steel) Long Gauge Block (Steel)	Gauge Block PTW B-1 GRADE: 00 Long Gauge Block TESA/---	In-house method: Gauge Block Calibration Procedure (Document No.: MT-C-95-014) In-house method: Long Gauge Block Calibration Procedure (MT-C-95-132)	125	mm	125	mm	Steel	0.50	μm
			150	mm	150	mm	Steel	0.52	μm
			175	mm	175	mm	Steel	0.56	μm
			200	mm	200	mm	Steel	0.58	μm
			250	mm	250	mm	Steel	0.66	μm
			300	mm	300	mm	Steel	0.74	μm
			400	mm	400	mm	Steel	0.90	μm
			500	mm	500	mm	Steel	1.08	μm
			0.5	mm	10	mm	Steel	0.10	μm
			> 10	mm	15	mm	Steel	0.11	μm
> 15	mm	25	mm	Steel	0.13	μm			
> 25	mm	75	mm	Steel	0.23	μm			
> 75	mm	100	mm	Steel	0.28	μm			
Approval Signatory: CHU, Wei-Hsin; WU, Min-Chao; HSU, Cherng-Chin; KAO, Wei-Chuan									
KA1006 Thickness gauge	Outside Micrometer /Mitutoyo /MDC-1"SB	In-house method: Calibration procedure for Thickness gauge (Document No.: MT-C-99-007)	0	mm	3	mm		0.002	mm
Approval Signatory: CHU, Wei-Hsin; WU, Min-Chao; HSU, Cherng-Chin; KAO, Wei-Chuan									



calibration items	working standard brand /model	calibration method document name /no.	measurand level or range			measurement conditions /independent variable explanation	smallest uncertainty	
			minimum value	units	maximum value		value	units
KA1017 3-wire gauge /pin gauge	Pin Gauge /MTC/---	In-house method: Calibration Procedure for Three Wire Units /Pin Gauge (Document No.: MT-C-95-134)	0.1	mm	15	mm	3-wire gauge: (Steel and ceramic)	1.1 $\mu$ m
Approval Signatory: CHU, Wei-Hsin; WU, Min-Chao; HSU, Chemng-Chin; KAO, Wei-Chuan								
KA2001 Rule	Standard Tape /B-Y/5 m	In-house method: Calibration Procedure for Rule (Document No.: MT-C-97-002)	0	cm	200	cm		0.04 cm
Approval Signatory: CHU, Wei-Hsin; WU, Min-Chao; HSU, Chemng-Chin; KAO, Wei-Chuan								
KA2002 Tape	Standard Tape /B-Y/5 m	In-house method: Calibration Procedure for Tape (Document No.: MT-C-95-095)	0	cm	500	cm		0.04 cm
Approval Signatory: CHU, Wei-Hsin; WU, Min-Chao; HSU, Chemng-Chin; KAO, Wei-Chuan								
KA2015 Thickness Gauge	Gauge Blocks /TSUGAMI/B1	In-house method: The Evaluation Report for Dial or Digital Thickness Gauge /MT-C-96-013	0	mm	10	mm	digital (resolution: 0.01 mm)	0.01 mm
Approval Signatory: CHU, Wei-Hsin; WU, Min-Chao; HSU, Chemng-Chin; KAO, Wei-Chuan								
KA2099 Measuring Wheel	Standard Tape	In-house method: Measuring Wheel Calibration Procedure (Document No.: MT-C-109-012)	0	m	1	m		0.02 m
Approval Signatory: CHU, Wei-Hsin; WU, Min-Chao; HSU, Chemng-Chin; KAO, Wei-Chuan								



Mass/Force calibration items	working standard brand /model	calibration method document name /no.	measurand level or range			measurement conditions /independent variable	smallest uncertainty		
			minimum value	units	maximum value		units	value	units
KC2004 Push Pull Gauge	STANDARD WEIGHTS METTLER /1 g to 200 g CHINA SCALES /100 g to 2000 g CHINA SCALES /100 g to 10 kg	In-house method: Calibration procedure for Push Pull Gauge (MT-C-106-015)	0.01 (0.001)	N (kgf)	9.8 (1)	N (kgf)	0.06 (0.006)	N (kgf)	
			0.01 (0.001)	N (kgf)	98 (10)	N (kgf)	0.10 (0.010)	N (kgf)	
			0.1 (0.01)	N (kgf)	490 (50)	N (kgf)	0.4 (0.04)	N (kgf)	
			0.5 (0.005)	N (kgf)	9.8 (1)	N (kgf)	0.10 (0.010)	N (kgf)	
			0.5 (0.05)	N (kgf)	98 (10)	N (kgf)	0.5 (0.05)	N (kgf)	
			2.5 (0.25)	N (kgf)	490 (50)	N (kgf)	1 (0.10)	N (kgf)	
Approval Signatory: CHU, Wei-Hsin; WU, Min-Chao; HSU, Cherng-Chin; KAO, Wei-Chuan									
KC4001 Torque Wrench	Torque Calibrator (NORBAR/50592.LOG/43228) Torque Calibrator (NORBAR/50593.LOG/43228) Torque Calibrator (NORBAR/50772.LOG/43228)	In-house method: Calibration Procedure for Torque Wrench (MT-C-110-004)	1	N·m	<25	N·m	CW, CCW	1.3	%
			25	N·m	<100	N·m	CW, CCW	1.7	%
			100	N·m	1000	N·m	CW, CCW	1.1	%
Approval Signatory: CHU, Wei-Hsin; HSU, Cherng-Chin; KAO, Wei-Chuan; HUANG, Chien-Ming									



calibration items	working standard brand /model	calibration method document name /no.	measurand level or range			measurement conditions /independent variable explanation	smallest uncertainty	
			minimum value	units	maximum value		value	units
KC4002 Torque Screwdriver	Torque Calibrator (Norbar/43213) Torque Calibrator (TOHNICHI /TDT60CN3-G)	In-house method: Calibration Procedure for Torque Driver (MT-C-110-005)	0.02	N·m	<0.6	CW, CCW	1.6	%
			0.6	N·m	10	CW, CCW	1.8	%
Approval Signatory: CHU, Wei-Hsin; HSU, Chermg-Chin; KAO, Wei-Chuan; HUANG, Chien-Ming								

Temperature/Humidity

calibration items	working standard brand /model	calibration method document name /no.	measurand level or range			measurement conditions /independent variable explanation	smallest uncertainty	
			minimum value	units	maximum value		value	units
KE1001 Liquid-in-glass Thermometer	Platinum Resistance Thermometer FLUKE/5681; Super Thermometer: FLUKE/1595A	In-house method: Liquid-in-Glass Thermometer Measurement System Calibration Procedure (Document No.: MT-C-95-015)	-80	°C	300	Total immersion & partial immersion	0.016	°C
Approval Signatory: WANG, Jian-Wha; CHU, Wei-Hsin; LIN, Chin-Liang; HSU, Chermg-Chin								





calibration items	working standard brand /model	calibration method document name /no.	measurand level or range			measurement conditions /independent variable explanation	smallest uncertainty	
			minimum value	units	maximum value		units	value
KE1004 THERMOCOUPLE (R, S, K, N, E, T, J type)	Platinum Resistance Thermometer FLUKE 5628 Platinum Resistance Thermometer FLUKE 5624 THERMOCOUPLE THERMOWAY /R TYPE	In-house method: Thermocouple Calibration Procedure (MT-C-107-008)	> 50	°C	≤ 600	E TYPE	0.13	°C
			> 600	°C	≤ 900	E TYPE	0.82	°C
			≥ -80	°C	≤ 50	J TYPE	0.22	°C
			> 50	°C	≤ 600	J TYPE	0.11	°C
			> 600	°C	≤ 700	J TYPE	0.80	°C
			≥ -80	°C	≤ 50	T TYPE	0.22	°C
			> 50	°C	≤ 400	T TYPE	0.11	°C
Approval Signatory: WANG, Jian-Wha; CHU, Wei-Hsin; LIN, Chin-Liang; HSU, Cherng-Chin; KAO, Wei-Chuan; TSAI, Shun-Chin								
KE1005 Thermocouple Thermometer	Standard Platinum Resistance Thermometer FLUKE/5624; FLUKE/5628; Thermocouple Thermometer THERMOWAY /R TYPE	In-house method: Thermocouple Thermometer Calibration Procedure (Document No.: MT-C-107-007)	≥ 0	°C	≤ 50	R Type	0.11	°C
			> 50	°C	≤ 600	R Type	0.11	°C
			> 600	°C	≤ 1000	R Type	0.83	°C
			> 1000	°C	≤ 1200	R Type	2.1	°C
			≥ 0	°C	≤ 50	S Type	0.11	°C
			> 50	°C	≤ 600	S Type	0.11	°C
			> 600	°C	≤ 1000	S Type	0.89	°C
			> 1000	°C	≤ 1200	S Type	2.1	°C
			≥ -80	°C	≤ 50	K Type	0.22	°C
			> 50	°C	≤ 600	K Type	0.16	°C
			> 600	°C	≤ 1000	K Type	0.83	°C
			> 1000	°C	≤ 1200	K Type	2.1	°C
			≥ -80	°C	≤ 50	N Type	0.22	°C
> 50	°C	≤ 600	N Type	0.13	°C			
> 600	°C	≤ 1000	N Type	0.83	°C			
> 1000	°C	≤ 1200	N Type	2.1	°C			
≥ -80	°C	≤ 50	E Type	0.22	°C			



calibration items	working standard brand /model	calibration method document name /no.	measurand level or range				measurement conditions /independent variable explanation	smallest uncertainty	
			minimum value	units	maximum value	units		value	units
KE1005 Thermocouple Thermometer	Standard Platinum Resistance Thermometer FLUKE/5624; FLUKE/5628; Thermocouple Thermometer THERMOWAY /R TYPE	In-house method: Thermocouple Thermometer Calibration Procedure (Document No.: MT-C-107-007)	> 50	°C	≤ 600	°C	E Type	0.13	°C
			> 600	°C	≤ 900	°C	E Type	0.82	°C
			≥ -80	°C	≤ 50	°C	J Type	0.22	°C
			> 50	°C	≤ 600	°C	J Type	0.11	°C
			> 600	°C	≤ 700	°C	J Type	0.80	°C
			≥ -80	°C	≤ 50	°C	T Type	0.22	°C
> 50	°C	≤ 400	°C	T Type	0.11	°C			
Approval Signatory: WANG, Jian-Wha; CHU, Wei-Hsin; LIN, Chin-Liang; HSU, Cherng-Chin; KAO, Wei-Chuan									
KE1010 temperature Chamber	Hydra Data Bucket FLUKE/1586A	In-house method: Temperature Humidity Chamber Calibration Procedure /MT-C-108-005	15	°C	50	°C		0.6	°C
			Approval Signatory: CHU, Wei-Hsin; HSU, Cherng-Chin; KAO, Wei-Chuan; TSAI, Shun-Chin						
KE1099 Portable Thermometer Data Logger	Hydra Data Bucket FLUKE/1586A	In-house method: Humidity/temperature instrument Measurement System Validation Program /MT-C-109-016	15	°C	50	°C		0.8	°C
			Approval Signatory: WANG, Jian-Wha; CHU, Wei-Hsin; HSU, Cherng-Chin; KAO, Wei-Chuan; TSAI, Shun-Chin						





calibration items	working standard brand /model	calibration method document name /no.	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
			minimum value	units	maximum value	units		value	units
KE1099 Temperature- Calibrator Furnace; Dry-Well Calibrator; High Temperature Furnace.	Platinum Resistance Thermometer FLUKE/5624; FLUKE/5628	In-house method: Temperature Calibration Furnace Measurement Calibration Procedure (Document No.: MT-C-103-022)	$\geq -80$	$^{\circ}\text{C}$	$\leq 600$	$^{\circ}\text{C}$		0.5	$^{\circ}\text{C}$
Approval Signatory: WANG, Jian-Wha; CHU, Wei-Hsin; LIN, Chin-Liang; HSU, Cherng-Chin									
KE2001 Hygrometer Hydrograph	Digital Hygrometer ROTRONIC /HYGROLOG HL- NT3-DP Hydra Data Bucket FLUKE/1586A	In-house method: Humidity /temperature instrument Measurement System Validation Program /MT-C-109-016	15	$^{\circ}\text{C}$	50	$^{\circ}\text{C}$		0.8	$^{\circ}\text{C}$
Approval Signatory: WANG, Jian-Wha; CHU, Wei-Hsin; HSU, Cherng-Chin; KAO, Wei-Chuan; TSAI, Shun-Chin									
KE2004 Temperature & Humidity recorder	Digital Hygrometer ROTRONIC /HYGROLOG HL- NT3-DP Hydra Data Bucket FLUKE/1586A	In-house method: Humidity /temperature instrument Measurement System Validation Program /MT-C-109-016	15	$^{\circ}\text{C}$	50	$^{\circ}\text{C}$		1.0	$^{\circ}\text{C}$
Approval Signatory: WANG, Jian-Wha; CHU, Wei-Hsin; HSU, Cherng-Chin; KAO, Wei-Chuan; TSAI, Shun-Chin									
KE2005 Temp. /Humidity Chamber	Digital Hygrometer ROTRONIC/HYGR OLOG HL-NT3-DP Hydra Data Bucket FLUKE/1586A	In-house method: Temperature Humidity Chamber Calibration Procedure /MT-C-108-005	15	$^{\circ}\text{C}$	50	$^{\circ}\text{C}$		0.6	$^{\circ}\text{C}$
Approval Signatory: WANG, Jian-Wha; CHU, Wei-Hsin; HSU, Cherng-Chin; KAO, Wei-Chuan; TSAI, Shun-Chin									



Electricity calibration items	working standard brand /model	calibration method document name /no.	measurand level or range			measurement conditions /independent variable	smallest uncertainty	
			minimum value	units	maximum value		units	value
KF1001 DCV source DC Voltmeter	CALIBRATOR DATRON4808 DC VOLTAGE MULTIMETER HP3458	In-house method: DC VOLTAGE MEASURING SYSTEM CALIBRATION PROCEDURE (Document No.: MT-C-95-010)	100	mV	1000	V	36	$\mu$ V/V
Approval Signatory: CHU, Wei-Hsin; LIN, Chin-Liang; SHE, Yen-Fu; HSU, Cherrig-Chin; JHAN, Yuan-Jhu								
KF1002 DCA source DC Current-meter	MULTIFUNCTION CALIBRATOR DATRON4808 DIGITAL MULTIMETER HP 3458A	In-house method: DC CURRENT MEASURING SYSTEM CALIBRATION PROCEDURE (Document No.: MT-C-95-012)	100	$\mu$ A	1	A	0.44	mA/A
Approval Signatory: CHU, Wei-Hsin; LIN, Chin-Liang; SHE, Yen-Fu; HSU, Cherrig-Chin; JHAN, Yuan-Jhu								
KF1003 DC High Voltage meter DC High Voltage tester	High Voltage DIVIDER /ROSS/V/D /60-12.5Y-B-KB-A DMM/FLUKE/87V	In-house method: DC high voltage measuring system calibration procedure (Document No.: MT-C-100-001)	1	kV	20	kV	2.0	%
Approval Signatory: CHU, Wei-Hsin; LIN, Chin-Liang; SHE, Yen-Fu; HSU, Cherrig-Chin; JHAN, Yuan-Jhu								



calibration items	working standard brand /model	calibration method document name /no.	measurand level or range			measurement conditions /independent variable	smallest uncertainty	
			minimum value	units	maximum value		units	value
KF1011 ACV source AC Voltmeter	Multifunction calibrator DATRON 4808 Digital multimeter HP 3458A	In-house method: AC voltage measuring system calibration procedure (Document No.: MT-C-95-011)	1	V	1000	V	0.50	mV/V
Approval Signatory: CHU, Wei-Hsin; LIN, Chin-Liang; SHE, Yen-Fu; HSU, Cherng-Chin; JHAN, Yuan-Jhu								
KF1012 ACA source AC Current-meter	Multifunction calibrator DATRON 4808 Digital multimeter HP 3458A	In-house method: AC current measuring system calibration procedure (Document No.: MT-C-95-013)	1	mA	1	A	2.9	mA/A
Approval Signatory: CHU, Wei-Hsin; LIN, Chin-Liang; SHE, Yen-Fu; HSU, Cherng-Chin; JHAN, Yuan-Jhu								
KF1013 AC High Voltage METER AC High Voltage tester	DIVIDER/ROSS/VD /60-12.5Y-B-KB-A DMM /FLUKE/87V	In-house method: AC high voltage measuring system calibration procedure (Document No.: MT-C-100-002)	1	kV	20	kV	2.0	%
Approval Signatory: CHU, Wei-Hsin; LIN, Chin-Liang; SHE, Yen-Fu; HSU, Cherng-Chin; JHAN, Yuan-Jhu								



calibration items	working standard brand /model	calibration method document name /no.	measurand level or range			measurement conditions /independent variable	smallest uncertainty		
			minimum value	units	maximum value		value	units	
KF3001 DC RESISTANCE METER	PRECISION DECADE RESISTORS IET HARS-X-3-001	In-house method: DC resistance measuring system calibration procedure (Document No.: MT-C-103-021)	0.1	$\Omega$	<1	Resistance meter	2	m $\Omega$ / $\Omega$	
			100	$\Omega$	<100	Resistance meter	0.05	m $\Omega$ / $\Omega$	
DC RESISTANCE	DC voltage multi meter HP3458A PRECISION DECADE RESISTORS BIDDLE 72-6346-1 PRECISION DECADE RESISTORS IET 1433-33		1	M $\Omega$	100	M $\Omega$	Resistance meter	53	$\mu\Omega$ / $\Omega$
			0.001	$\Omega$	0.001	M $\Omega$	Resistance meter	0.52	m $\Omega$ / $\Omega$
			0.01	$\Omega$	0.01	$\Omega$	Resistance meter	0.02	$\Omega$ / $\Omega$
			1	$\Omega$	1	$\Omega$	Resistance meter	2	m $\Omega$ / $\Omega$
			10	$\Omega$	10	$\Omega$	Resistance meter	0.05	m $\Omega$ / $\Omega$
			100	$\Omega$	100	$\Omega$	Resistance meter	34	$\mu\Omega$ / $\Omega$
			1	k $\Omega$	1	k $\Omega$	Resistance meter	32	$\mu\Omega$ / $\Omega$
			10	k $\Omega$	10	k $\Omega$	Resistance meter	30	$\mu\Omega$ / $\Omega$
			100	k $\Omega$	100	k $\Omega$	Resistance meter	30	$\mu\Omega$ / $\Omega$
			1	M $\Omega$	1	M $\Omega$	Resistance meter	30	$\mu\Omega$ / $\Omega$
10	M $\Omega$	10	M $\Omega$	Resistance meter	55	$\mu\Omega$ / $\Omega$			
100	M $\Omega$	100	M $\Omega$	Resistance meter	89	$\mu\Omega$ / $\Omega$			
0.001	$\Omega$	<0.01	$\Omega$	Resistance meter	0.28	m $\Omega$ / $\Omega$			
0.01	$\Omega$	<1	$\Omega$	Resistance source	9	m $\Omega$ / $\Omega$			
1	$\Omega$	<100	$\Omega$	Resistance source	7.1	m $\Omega$ / $\Omega$			
100	$\Omega$	<1	M $\Omega$	Resistance source	81	$\mu\Omega$ / $\Omega$			
1	M $\Omega$	100	M $\Omega$	Resistance source	73	$\mu\Omega$ / $\Omega$			
1	$\Omega$	1	$\Omega$	Resistance source	0.6	m $\Omega$ / $\Omega$			
10	$\Omega$	10	$\Omega$	Resistance source	81	$\mu\Omega$ / $\Omega$			
100	$\Omega$	100	$\Omega$	Resistance source	71	$\mu\Omega$ / $\Omega$			
1	k $\Omega$	1	k $\Omega$	Resistance source	73	$\mu\Omega$ / $\Omega$			
10	k $\Omega$	10	k $\Omega$	Resistance source	48	$\mu\Omega$ / $\Omega$			
100	k $\Omega$	100	k $\Omega$	Resistance source	48	$\mu\Omega$ / $\Omega$			
1	M $\Omega$	1	M $\Omega$	Resistance source	49	$\mu\Omega$ / $\Omega$			
				Resistance source	0.6	m $\Omega$ / $\Omega$			



calibration items	working standard brand /model	calibration method document name /no.	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
			minimum value	units	maximum value	units		value	units
KF3001 DC RESISTANCE METER DC RESISTANCE	PRECISION DECADE RESISTORS IET HARS-X-3-001 DC voltage multi meter HP3458A PRECISION DECADE RESISTORS BIDDLE 72-6346-1 PRECISION DECADE RESISTORS IET 1433-33	In-house method: DC resistance measuring system calibration procedure (Document No.: MT-C-103-021)	10	MΩ	10	MΩ	Resistance source	0.6	mΩ/Ω
			100	MΩ	100	MΩ	Resistance source	0.6	mΩ/Ω
			0.001	Ω	<0.1	Ω	Resistance meter	0.02	Ω/Ω
			0.1	Ω	0.1	Ω	Resistance meter	0.3	mΩ/Ω
			0.001	Ω	0.001	Ω	Resistance source	9	mΩ/Ω
			0.01	Ω	0.01	Ω	Resistance source	7.1	mΩ/Ω
			0.1	Ω	0.1	Ω	Resistance source	0.3	msΩ/Ω
Approval Signatory: CHU, Wei-Hsin; LIN, Chin-Liang; SHE, Yen-Fu; HSU, Cherng-Chin; JHAN, Yuan-Jhu									
KF3008 Insulation Resistance Tester	Decade Standard Resistance Box BIDDLE/72-6346-1	In-house method: Digital Insulation Resistance meter /Insulation Resistance Calibration Procedure (MT-C-108-007)	1	MΩ	20	MΩ	100 V	20	msΩ/Ω
			20	MΩ	100	MΩ	100 V	0.2	Ω/Ω
			1	MΩ	20	MΩ	250 V	20	mΩ/Ω
			20	MΩ	200	MΩ	250 V	0.2	Ω/Ω
			1	MΩ	20	MΩ	500 V	20	mΩ/Ω
			20	MΩ	200	MΩ	500 V	0.2	Ω/Ω
			200	MΩ	500	MΩ	500 V	1	Ω/Ω
			1	MΩ	200	MΩ	1000 V	0.2	Ω/Ω
			200	MΩ	2000	MΩ	1000 V	2	Ω/Ω
			2	GΩ	10	GΩ	1000 V	1	Ω/Ω
Approval Signatory: CHU, Wei-Hsin; LIN, Chin-Liang; SHE, Yen-Fu; HSU, Cherng-Chin; KAO, Wei-Chuan; JHAN, Yuan-Jhu									



calibration items	working standard brand /model	calibration method document name /no.	measurand level or range			measurement conditions /independent variable explanation	smallest uncertainty	
			minimum value	units	maximum value		value	units
KF5003 Electrostatic field meter	AC /DC High voltage divider /ROSS /VD60-12.5Y-B-KB-AL	In-house method: Calibration procedure for electrostatic field meter (Document No.: MT-C-95-080)	1	kV	20	kV	0.08	kV/kV
Approval Signatory: CHU, Wei-Hsin; LIN, Chin-Liang; SHE, Yen-Fu; HSU, Cherng-Chin; JHAN, Yuan-Jhu								

Electromagnetics

calibration items	working standard brand /model	calibration method document name /no.	measurand level or range			measurement conditions /independent variable explanation	smallest uncertainty	
			minimum value	units	maximum value		value	units
KG3001 Illuminance Meter	Standard Lamp ORIEL 63350 71171 Standard Lamp OL 100U HTS-360 Ipicoammeter /detector KEITHLEY /LMT 6485 /--- 4459256/11B7251 dc powersupply Newport 69935 234	In-house method: Illuminance Meter Calibration Procedure (Document No.: MT-C-95-151)	≥ 50	lx	< 500	lx	1.2	%
			≥ 500	lx	< 1000	lx	1.4	%
			≥ 1000	lx	< 1500	lx	1.2	%
			≥ 1500	lx	< 5000	lx	1.2	%
			≥ 5000	lx	< 10000	lx	1.2	%
			≥ 10000	lx	< 15000	lx	1.5	%
			≥ 15000	lx	< 20000	lx	1.5	%
			= 20000	lx		lx	1.5	%
Approval Signatory: WANG, Jian-Wha; CHU, Wei-Hsin; SHE, Yen-Fu; HSU, Cherng-Chin								



Chemical calibration items	working standard brand /model	calibration method document name /no.	measurand level or range			measurement conditions /independent variable	smallest uncertainty	
			minimum value	units	maximum value		value	units
K12000 pH meter	1. Multifunction calibrator: AOP/PJ6301/351004 2. pH Meter: METTLER /S475-Basic/B243450359 3. pH standard buffer / (4, 7, 10) pH	In-house method: pH meter Calibration Procedure (Document No.: MT-C-95-149)	pH 1		pH 1	@ 25 °C	0.02	pH
			pH 4		pH 4	@ 25 °C	0.01	pH
			pH 7		pH 7	@ 25 °C	0.01	pH
			pH 10		pH 10	@ 25 °C	0.01	pH
			pH 13		pH 13	@ 25 °C	0.02	pH
			20	°C	20	Temperature Compensations (slope)	0.30	%
			60	°C	60	Temperature Compensations (slope)	0.40	%
						Electrode Slope (@ 25 °C)	1.5	%
			pH 4		pH 4	Calibration Curve (@ 25 °C)	0.10	pH
			pH 7		pH 7	Calibration Curve (@ 25 °C)	0.10	pH
pH 10		pH 10	Calibration Curve (@ 25 °C)	0.10	pH			
Approval Signatory: CHU, Wei-Hsin; HSU, Cherrng-Chin; TSAI, Shun-Chin								



calibration items	working standard brand /model	calibration method document name /no.	measurand level or range			measurement conditions /independent variable		smallest uncertainty	
			minimum value	units	maximum value	units	explanation	value	units
KI9900 Conductivity Meter	1. Conductivity Meter: METTLER /S475-Basic /B243450359	In-house method: Calibration Process of Electric Conductivity meter (Document No.: MT-C-95-158)	84	$\mu\text{S/cm}$	84	$\mu\text{S/cm}$		3.0	%
			1413	$\mu\text{S/cm}$	1413	$\mu\text{S/cm}$		2.0	%
	2. Conductivity Standard Solution /84, 1413, 12880 $\mu\text{S/cm}$		12880	$\mu\text{S/cm}$	12880	$\mu\text{S/cm}$		2.0	%
Approval Signatory: CHU, Wei-Hsin; HSU, Cherng-Chin; TSAI, Shun-Chin									

Note: Smallest uncertainty represents an expanded uncertainty using a coverage factor approximately 95 % level of confidence. (Null Below)

